




# Programming Note

8566A/8568A/9835/9845-1

NOVEMBER 1980

SUPERCEDES: NONE

## Introductory Operating Guide for the 8566A/8568A Spectrum Analyzers with the 9835/9845 Desktop Computers



© HEWLETT-PACKARD CO. 1980



# **HP MANUAL SCANS**

**By**

***Artek Media***

1042 Plummer Cir. SW  
Rochester, MN 55902

[www.artekmedia.com](http://www.artekmedia.com)

"High resolution scans of obsolete technical manuals"

**ALL HP MANUALS ARE REPRODUCED  
WITH PERMISSION OF  
AGILENT TECHNOLOGIES INC.**

**Reproduction of this scanned document or removal of this disclaimer  
Will be considered in violation of both Agilent's and Artek Media's copy rights**


If your looking for a quality scanned technical manual in PDF format please visit our WEB site at [www.artekmedia.com](http://www.artekmedia.com) or drop us an email at [manuals@artekmedia.com](mailto:manuals@artekmedia.com) and we will be happy to email you a current list of the manuals we have available.

If you don't see the manual you need on the list drop us a line anyway we may still be able to point you to other sources. If you have an existing manual you would like scanned please write for details, This can often be done very reasonably in consideration for adding your manual to our library.

Typically the scans in our manuals are done as follows;

- 1) Typed text pages are typically scanned in black and white at 300 dpi.
- 2) Photo pages are typically scanned in gray scale mode at 600 dpi
- 3) Schematic diagram pages are typically scanned in black and white at 600 dpi unless the original manual had colored high lighting (as is the case for some 70's vintage Tektronix manuals).

Thanks



Dave & Lynn Henderson  
Artek Media

## Introduction

This note is an introductory guide to remote operation and programming of 8566A and 8568A Spectrum Analyzers using either the 9835 or 9845 Desktop Computer. Included in this guide are system connections for remote operation and several example programs with descriptions of each step.

The 8566A and 8568A are microprocessor-controlled, general purpose spectrum analyzers which are compatible with the Hewlett-Packard Interface Bus (HP-IB). When used with any HP-IB controller, such as the 9835 or 9845, these instruments become fully automated spectrum analyzers featuring:

- \* Precise, stable LO tuning
- \* High sensitivity and resolution
- \* Wide dynamic range

## Related Documents

Complete operating information for the 8566A/8568A analyzers can be found in:

1. 8566A/8568A Spectrum Analyzer Operation  
(P/N 08566-90002 or 08568-90002)
2. 8566A/8568A Spectrum Analyzer Remote Operation  
(P/N 08566-90003 or 08568-90003)
3. 8566A/8568A Spectrum Analyzer Pull-Out Information Cards

Information on operating the 9835 and 9845 controllers can be found in:

1. System 35/45 Operating and Programming Manual
2. System 35/45 Beginner's Guide
3. System 35/45 Reference Guide
4. System 35/45 I/O ROM Programming

A description of interface programming and hardware can be found in:

1. BASIC Language Interfacing Concepts  
(P/N 09835-90600)

## Equipment Required

To perform the examples in this note, you will need the following equipment and accessories:

1. 8566A or 8568A Spectrum Analyzer
2. 9835A/B Desktop Computer with 98332A I/O ROM, or  
9845B/T Desktop Computer with 98412A I/O ROM (Option 312)
3. 98034A HP-IB Interface\*

## Setup

Figure 1 shows the system connections and switch setting for the 98034A HP-IB Interface. To connect the system as shown, follow these steps:

\*revised cards only

1. Turn off power to the 9835/9845.
2. Install the I/O ROM in any available socket (in front on the 9835, or in the left side drawer on the 9845).
3. Install the 98034A in any available socket on the rear of the 9835/9845. Be sure the 98034A seats securely in the socket; this has occurred when the latch on top of the interface pops up, locking the card into the socket.
4. Set the rotary switch located on top of the 98034A to position 7. Seven is the select code of the interface for all programs in this guide.
5. Connect the 24-pin connector at the free end of the 98034A cable to the rear panel of the 8566A/8568A (see Figure 1). The connector is shaped to ensure proper orientation.

### CAUTION

**Do not attempt to mate silver English threaded screws on one connector with black metric threaded nuts on another connector, or vice versa, as damage to the hardware may result. A metric conversion kit which will convert one cable and one or two instruments to metric hardware may be obtained by ordering HP P/N 5060-0138.**





Figure 1. System connection.


## Check-Out

After making AC power line connections to the analyzer, the STANDBY lights on both the RF and display sections should be illuminated. Switch on the 9835/9845 and set the analyzer to LINE ON.



Upon LINE ON, the analyzer will perform an automatic internal instrument check, designated by the red INSTR CHECK indicators. Both LED's will turn on momentarily during the brief check routine and, if the instrument is operating properly, will go off and remain off during operation, except when another instrument check is triggered by an Instrument Preset. If one or both LED's remain on, refer to the 8566A/8568A Operating and Service Manual, Section II.

Verify that the analyzer's address is set to 18. The read/write address of the 8566A or 8568A can be determined and altered from the front panel by using the shift function P:



sets the address to 18.

When the analyzer is turned on from a cold state, crt messages OVEN COLD and REF UNLOCK may appear. These will go off typically ten minutes after AC power is connected. Type the following commands on the controller keyboard:

**ABORT10 7** (Press EXECUTE)  
**REMOTE 718** (Press EXECUTE).

If ADRS'D and REM light up on the analyzer's front panel, proceed to the programming examples. If either ADRS'D or REM do not light, check to make sure that the 98034A select code is set to 7, the interface cables are properly connected, and the address in the REMOTE statement matches the address of the 8566A/8568A. Although 18 is the factory-set address and the address used in the following examples, other addresses are possible.

If both ADRS'D and REM still do not light, consult the 8566A/8568A Operating and Service Manual, the System 35/45 Test Manual, and the 98034A Installation and Service Manual for troubleshooting information.


### Programming Examples

The following examples illustrate some of the ways to operate the 8566A/8568A using the 9835/9845 controller.


The examples illustrate setting front panel controls remotely and outputting their values, outputting marker values, and outputting trace data. An example harmonic distortion measurement program incorporates some of these techniques in a typical application.

#### EXAMPLE 1: PROGRAMMING FRONT-PANEL FUNCTIONS

To preset the analyzer, and set center frequency to 100 MHz and span to 10 MHz, enter the following on the keyboard of the 9835/9845 controller: `OUTPUT 7 18; "IP CF 100MZ SP 10MZ"`



Executing this statement initiates the sequence of operations shown above. The final crt display with a 100 MHz signal present should look like this:



The last function activated, SPAN, will appear with its current value on the analyzer crt as shown in the shaded box.

**NOTE**

An important concept in analyzer programming is worthy of special note here. The sequence of operations executed above could have been entered manually from the front panel of the analyzer to yield the same result. In fact, a manual sequence of keystrokes is usually developed first and then used as a basis for executing the same procedure under program control. This simple technique is recommended as a powerful tool for software development with the automatic spectrum analyzer.

**EXAMPLE 2: OUTPUTTING A FUNCTION OR MARKER VALUE**

In the first case, a BASIC program is shown which directs the analyzer to activate center frequency, and to prepare to output the current value in a subsequent statement. The value is then transferred into the variable F and printed. The END statement, line 50, terminates the program.

```

10  OUTPUT 718;"CF OR"           ! Activate center frequency, prepare
20                               ! to output value of active function.
30  ENTER 718;F                 ! Transfer value to F.
40  PRINT "Center Frequency =";F;"Hz" ! Print value.
50  END

```

To enter the program, press:

EDIT [Press EXECUTE].

10 \_

should appear. Type a line and press STORE. Now

20 \_

should appear. Continue entering program code line by line. After storing the last line, END, press RUN to execute the program.\* (Omit annotation which begins with "!" on each line, or entire lines which contain only annotation; these comments are provided for the reader's clarification only. Note that your line numbers will not in general correspond to those in this guide.)

A typical output would be:

Center Frequency = 100000000 Hz

Next, we would like to output both the amplitude and frequency of the active marker. To illustrate this, connect the analyzer's CAL OUTPUT to the RF INPUT. Type SCRATCH A and press EXECUTE to clear the program memory, and enter the following program:

```

10  OUTPUT 718;"IP FA75MZ FB150MZ S2 TS E1"
20                               ! Instrument preset, set start and stop freq's,
30                               ! single sweep, take sweep, peak search.
40  OUTPUT 718;"MA"             ! Prepare to output marker amplitude.
50  ENTER 718;A                 ! Transfer amplitude into variable A.
60  OUTPUT 718;"MF"            ! Prepare to output marker frequency.
70  ENTER 718;F                 ! Transfer frequency into variable F.
80  PRINT A;"dBm    ";F/1E6;"MHz" ! Print A and F (scaled to megahertz).
90  END

```

\*For a brief introduction to the controller editing facilities, refer to the Editing Section of the chapter entitled Keyboard Operations in the System 35/45 Operating and Programming Manual.

The first line presets the analyzer, sets start and stop frequencies to 75 MHz and 150 MHz, and then instructs the analyzer to use the single sweep mode. To ensure that a trace is displayed which corresponds to the current instrument control settings, a take sweep command ("TS") is used. This triggers a sweep and prevents the analyzer from accepting further commands until the trace is complete.

Upon completion of this sweep, the peak search ("E1") command is invoked, placing a marker on the largest signal displayed. Lines 40 and 50 instruct the analyzer to output the amplitude value in dBm into the variable A, and lines 60 and 70 cause the frequency value in hertz to be transferred into F. These two values are then printed with appropriate units. Note that the frequency in hertz has been divided by one million to yield megahertz.

Pressing RUN yields typical output:

-10.4 dBm      100.2 MHz

### EXAMPLE 3: OUTPUT TRACE DATA

An important capability of an automatic spectrum analyzer is to transfer trace amplitude data into an array in the controller for subsequent manipulation. A direct approach is shown in the first program:

```

10  DIM A(1000)           ! Dimension array A from 0 to 1000
20                          ! (1001 points total).
30  OUTPUT 718;"S2 TS 03 TA" ! Using 03 format (reference level units),
40                          ! prepare to output trace A.
50  FOR N=0 TO 1000      ! Begin FOR-NEXT loop.
60  ENTER 718;A(N)       ! Transfer formatted data one point at a time
70                          ! into A array.
80  NEXT N               ! End of FOR-NEXT loop.
90  FOR N=0 TO 1000 STEP 100 !
100 PRINT N,A(N)         ! Print every one-hundredth point
110 NEXT N               !
120 END

```

After dimensioning the array, four commands are sent to the analyzer in the OUTPUT 718 statement. First, the analyzer is set to the single sweep mode, followed by a take sweep command. The single sweep mode ("S2") is especially important when outputting trace data because it provides a static display while the values are being accessed. Following the TS command (discussed in Example 2) there is an output format command 03. (This is the letter O for Output, not zero!) The analyzer in this mode scales the display units from the ADC (analog-to-digital converter) to reference level units (in this example, dBm), and re-formats these values into a sequence of ASCII characters which will be transmitted over the interface bus. TA specifies trace A data, which are subsequently transferred one point at a time into the A array using the ENTER 718 statement repeated 1001 times.

Finally, to show what has happened, several data values are printed.

```

0           -91.9
100        -85.2
200        -86.4
300        -85.5
400        -82.1
500        -10.9
600        -77
700        -81.8
800        -87.8
900        -84.3
1000       -87

```



The running time for this program using a 9835 is about 26 seconds. To achieve a faster transfer, we can avoid rescaling the ADC values and re-formatting into ASCII code by using O2 instead of O3 output format. We can then achieve a very efficient means of transferring the trace data as unformatted binary values through the use of a byte-by-byte fast handshake command.

In the case below, a sequence of 8-bit bytes is transferred into the integer-valued A array. Note that the values in the A array are two bytes or sixteen bits long, as are the binary values to be transferred from the analyzer in the O2 format mode. Therefore, we have specified that 2002 bytes be transferred, which corresponds exactly to 1001 values. The values which are printed from the A array are in display units. These range from 0 to 1023, and may be accessed as such for further processing. A typical execution time for this transfer using the 9835/9845 is 200 milliseconds.

```

10  INTEGER A(1000)                ! Dimension A array from 0 to 1000
20                                  ! (1001 points total).
30  OUTPUT 718;"S2 TS O2 TA"      ! Single sweep, take sweep, using format O2
40                                  ! (binary units) prepare to output trace A.
50  ENTER 718 BFHS 2002 NOFORMAT;A(*) ! Transfer data into array using byte-
60                                  ! by-byte fast handshake, no format.
70  FOR N=0 TO 1000 STEP 100      !
80  PRINT N,A(N)                  ! Print every one-hundredth point.
90  NEXT N                          !
100 END

```

0	108
100	133
200	126
300	124
400	186
500	890
600	249
700	119
800	153
900	149
1000	125

This program illustrates how more advanced BASIC programming techniques can be implemented to produce significantly higher performance in the area of automatic instrument control. Such topics as advanced transfer techniques are treated in the System 35/45 I/O ROM Programming manuals.

#### NOTE

**Correct format usage when transferring data and commands to and from the analyzer is essential for proper operation under remote control. Errors in formatting are a frequent cause of program failure; study the format codes if you are not certain of correct usage when debugging a program under development.**

**Data are transferred over the interface bus one 8-bit byte at a time. These may be ASCII-encoded alphanumeric characters, or binary values. For example, when the O3 format has been specified (this is the default mode on instrument preset) and a trace value is output from the analyzer, a sequence of ASCII characters is transmitted across the bus, as many as needed to specify the value of interest. The analyzer automatically performs the necessary formatting from an internally stored binary value to an ASCII string, and the controller reverses this process on receipt of such a string. As the number of characters transferred is variable, a free field format is required in the control program.**

Alternatively, data values themselves may be transferred in 8-bit bytes (two bytes will be necessary to retain the full 10-bit precision of values stored in the analyzer). Here, the analyzer may be in the O2 format, and the controller in an unformatted or binary formatted mode (i.e., ASCII formatting must not occur). This is illustrated in the second trace output example involving the byte-by-byte fast hand shake transfer mode.

See the Spectrum Analyzer Remote Operation manual for further information on input/output formats.

#### EXAMPLE 4: HARMONIC DISTORTION MEASUREMENT

An example program which illustrates some of the techniques demonstrated above is included here. This program makes a harmonic distortion measurement by locating and measuring a signal's second and third harmonics and calculating the percent distortion relative to the fundamental. The technique suggested in Example 1 - converting a manual sequence of keystrokes into a program to perform the same functions - was used in developing the present example.

```

10  ! HARMONIC DISTORTION MEASUREMENT
20  ! REV A, 801024
30  !
40  OUTPUT 718;"IP"
50  LOCAL 718
60  DISP "Set analyzer to display the fundamental signal."
70  PAUSE
80  DISP ""
90  OUTPUT 718;"SP 03 0A"          ! Prepare to output the current span.
100 ENTER 718;Span                ! Transfer value (in hertz) to "Span".
110 Span=MIN(Span,1E5)           ! Use current value or 100 kHz,
120                               ! whichever is smaller.
130 OUTPUT 718;"S2 TS E1 MT1 SP",Span,"HZ TS MT0 E4 TS E1 E3 MA"
140                               ! Acquire signal with peak search, auto-zoom, marker to reference
150                               ! level, peak search; enter CF STEP SIZE with E3 command; use MA
160                               ! to prepare to output fundamental amplitude.
170 ENTER 718;Fund               ! Transfer marker amplitude to "Fund".
180 OUTPUT 718;"MF"              ! Prepare to output marker frequency.
190 ENTER 718;Freq               ! Transfer marker freq to "Freq"
200 Freq=Freq/1E6                ! Scale frequency to megahertz.
210 OUTPUT 718;"CF UP TS E1 MA"  ! Increment center freq by fundamental freq.
220 ENTER 718;Second            ! Transfer marker amplitude to "Second".
230 OUTPUT 718;"CF UP TS E1 MA"  ! Increment center freq by fundamental freq.
240 ENTER 718;Third             ! Transfer marker amplitude to "Third".
250 Dist=100*SQR(FNLin(Second)^2+FNLin(Third)^2)/FNLin(Fund)
260                               ! Compute root-sum-of-the-squares
265                               ! total harmonic distortion using "Lin"
270                               ! function defined below.
280                               !
290 Format1:  IMAGE 4A,XSDDD.DX,"dBm",XXXXX,"MHz" ! _____! _____! _____!
300 PRINT USING Format1;"Fund",Fund,Freq          ! _____! _____! Formatted
310 Format2:  IMAGE 2(4A,XSDDD.DX,"dBm",/ )       ! _____! _____! output.
320 PRINT USING Format2;"2nd ",Second,"3rd ",Third ! _____! _____!
330 PRINT USING "K,DDD.DD,K//";"Harmonic Distortion = ";Dist;"%" ! _____!
340                               !
350 DEF FNLin(X)=10^(X/20)           ! Function to compute linear value from dB's.
360                               !
370 END

```

Line 50 places the analyzer under front panel control allowing the operator to tune the analyzer to position the signal on screen. The span must be chosen such that the signal of interest is the largest response on the screen. When ready, the operator presses CONTInue. The program determines the present span and compares it to 100 kHz, choosing the smaller value. Then, a sweep is taken in single sweep mode, and peak search places the marker on the largest signal, i.e., the fundamental. Marker track is invoked to perform an Auto-Zoom to the span selected above. The signal is then moved to the reference level, the center frequency step size is set to the fundamental frequency, and the amplitude and frequency are output to the controller.

In line 210, the center frequency is incremented once to place the second harmonic on screen. Peak search locates the response and the marker amplitude is output. The same procedure is performed on the third harmonic in line 230.

In line 250, the percent distortion is computed as the root sum of the squares normalized to the fundamental amplitude. As linear values are required in this calculation, a function has been defined in line 350 which converts the dBm values to linear values. The results are finally printed according to the output formats in lines 290—330\*.

A typical harmonic distortion measurement might yield the following output:

```
Fund -10.1 dBm    100.0004 MHz
2nd  -41.6 dBm
3rd  -50.8 dBm

Harmonic Distortion =    2.82%
```

\*A discussion of PRINT and IMAGE statements can be found in the System 35/45 Operating and Programming Manual.

## 8568A PROGRAMMING CODE LIST

### FRONT PANEL COMMANDS

AT Input attenuation * A1 Clear-write trace A A2 Max Hold trace A A3 Store and view trace A A4 Store and blank trace A BL B - DL → B B1 Clear-write trace B B2 Max hold trace B B3 Store and view trace B * B4 Store and blank trace B * CA Coupled input attenuation CF Center frequency * CR Coupled resolution BW * CS Coupled step size * CT Coupled sweep time * CV Coupled video BW * C1 A - B off C2 A - B → A DB dB DL Display line DM dBm DN Step down DT Label terminator EE Enable number entry EK Enable DATA knob * EM Erase trace C memory EX Exchange A and B E1 Peak Search E2 Enter marker into center frequency E3 Enter marker/Δ frequency → step size E4 Enter marker amplitude → reference level * FA Start frequency * FB Stop frequency FS 0 - 1.5 GHz span	GZ GHz * HD Hold HZ Hz IP Instrument preset I1 Left RF input * I2 Right RF input KS Shift front panel keys * KSA Amplitude in dBm KSB Amplitude in dBmV KSC Amplitude in dBuV KSD Amplitude in voltage KSE Title KSF Measure sweep time KSG Video averaging on * KSH Video averaging off KSI Extended reference level range KSJ Manual DAC control KSK Count pilot IF at marker KSL Noise level off * KSM Noise level on KSN Count VTO at marker KSO Enter Δ - span KSP Set HP-IB address KSQ Count signal IF KSR Diagnostics on * KSS Second LO auto KST Second LO down KSU Second LO up KSV Frequency offset KSW Error correction routine KSX Use correction data KSY Do not use correction data KSZ Amplitude offset * KSa Normal detection	KSb Positive peak detection KSc A + B - A KSD Negative peak detection KSe Sample detection KSf Power on in last state KSG CRT beam off KSh CRT beam on KSi Exchange B and C KSj View trace C KSk Blank trace C KSl Trace B - trace C KSm Graticule blanked * KSn Graticule on KSo Characters blanked * KSp Characters on KSq Step gain off KSr Service request 102 KSt Continue sweep from marker KSu Stop at marker, single sweep KSV Inhibit phase lock KSw Display correction data normal EXT trigger KSx normal EXT trigger KSy normal VID trigger KSz Display storage address KS, Mixer level KS - Negative entry KS = Counter resolution KS( Save registers locked KS) Save registers unlocked KS > Preamp gain, input 2 KS < Preamp gain, input 1 KS  Display storage write KZ kHz * LG Enter log scale LN Linear scale	L0 Display line off * MC0 Marker frequency count off MC1 Marker frequency count on MS msec MV mV * MT0 Marker signal track off MT1 Marker signal track on MZ MHz * M1 Marker off M2 Marker normal M3 Marker Δ M4 Marker zoom RB Resolution BW RC Recall RL Reference level SC sec SP Frequency span SS Center frequency step size ST Sweep time SV Save * S1 Sweep continuous S2 Sweep single TH Enter threshold * T0 Threshold off * T1 Trigger free run T2 Trigger line T3 Trigger external T4 Trigger video UP Step up UR Upper right US μsec UV μV VB Video BW 0 to 9 . Decimal point or period
---	---	--	---

### OUTPUT COMMANDS

DR Read display and increment address EE Enable number entry KS123 <sup>10</sup> Output up to 1001 words LL Lower left recorder output MA Marker amplitude output	MF Marker frequency output OA Output active function OL Output learn string OT Output display text O1 Output format ASCII display units O2 Output format two 8 bit binary bytes	* O3 Output format ASCII parameter or instrument units O4 Output format one 8 bit binary byte TA Output trace A TB Output trace B UR Upper right recorder output
---	--	--

### DISPLAY INPUT COMMANDS

* DA Display address DD Display write DW Write into display and increment address * D1 Display size normal D2 Display size full CRT D3 Display size expand	GR Graph IB Input trace B, binary KS125 <sub>10</sub> Input up to 1001 display memory words LB Label PA Plot absolute * PD Pen down	PR Plot relative PS Skip to next display page PU Pen up SW Skip to next control instruction TS Take sweep
---	--	---

### SERVICE REQUEST COMMANDS

R1	R2	* R3	R4	SRQ	Command	Bit	Definition
Allow only SRQ 140	Allow SRQ 140 and 104	Allow SRQ 140 and 110	Allow SRQ 140 and 102	102	R4	1	units key pressed
				104	R2	2	end of sweep
				110	R3	3	hardware broken
				140	all	5	illegal command
				1xx	—	6	universal HP-IB service

\* selected with instrument preset

## 8566A PROGRAMMING CODE LIST

FRONT PANEL COMMANDS							
AT	Input attenuation	* HD	Hold	KSf	Power on in last state	* LG	Enter log scale
* A1	Clear-write trace A	HZ	Hz	KSg	CRT beam off	LL	Lower Left
A2	Max Hold trace A	IP	Instrument preset	KSj	CRT beam on	LN	Linear scale
A3	Store and view trace A	KS	Shift front panel keys	KSj	Exchange B and C	LO	Display line off
A4	Store and blank trace A	* KSA	Amplitude in dBm	KSj	View trace C	MS	msec
BL	B - DL → B	KSB	Amplitude in dBmV	KSj	Blank trace C	MV	mV
B1	Clear-write trace B	KSC	Amplitude in dBuV	KSj	Trace B → trace C	* MTO	Marker signal track off
B2	Max hold trace B	KSD	Amplitude in voltage	KSj	Graticule blanked	MT1	Marker signal track on
B3	Store and view trace B	KSE	Title	* KSn	Graticule on	MZ	MHz
* B4	Store and blank trace B	KSF	Measure sweep time	KSj	Characters blanked	* M1	Marker off
* CA	Coupled input attenuation	KSG	Video averaging on	* KSp	Characters on	M2	Marker normal
CF	Center frequency	* KSH	Video averaging off	KSq	Step gain off	M3	Marker Δ
* CR	Coupled resolution BW	KSj	Extended reference level range	KSr	Service request 102	M4	Marker zoom
* CS	Coupled step size	KSJ	DAC control	KSj	Band lock	PP	Preselector peak
* CT	Coupled sweep time	KSK	Marker to next peak	KSu	Stop at marker, single sweep	RB	Resolution BW
* CV	Coupled video BW	KSL	Noise level off	KSv	Signal identifier ext mixer	RC	Recall
* C1	A - B off	KSM	Noise level on	KSv	Display correction data normal EXT trigger	RL	Reference level sec
C2	A - B → A	KSN	Marker to minimum	KSx	normal EXT trigger	SC	sec
DB	dB	KSO	Enter Δ - span	KSy	normal VID trigger	SP	Frequency span
DL	Display line	KSP	Set HP-IB address	KSz	Display storage address	SS	CF step size
DM	dBm	KSQ	Band unlock	KS.	Mixer level	ST	Sweep time
DN	Step down	KSR	Diagnostics on	KS#	Turns off YTX self-heating correction	SV	Save
DT	Label terminator	KSS	Fast HP-IB	KS =	Factory preselector setting	* S1	Sweep continuous
EE	Enable number entry	KST	Fast preset 2 — 22 GHz	KS -	Negative entry	S2	Sweep single
EK	Enable DATA knob	KSU	External mixer preset	KS(	Save registers locked	TH	Enter threshold
* EM	Erase trace C memory	KSV	Frequency offset	KS)	Save registers unlocked	* T0	Threshold off
EX	Exchange A and B	KSW	Error correction routine	KS!	Display storage write	* T1	Trigger free run
E1	Peak Search	KSX	Use correction data	KS!	Display storage write	T2	Trigger line
E2	Enter marker into center frequency	KSY	Do not use correction data	KS#	Turns off YTX self-heating correction	T3	Trigger external
E3	Enter marker/Δ frequency — step size	KSZ	Amplitude offset	KS/	Manual Preselector peak	T4	Trigger video
E4	Enter marker amplitude → reference level	* KSA	Normal detection	KS< 9 2 >	Enter DL, TH, M2, M3 in display units	UP	Step up
* FA	Start frequency	KSc	Positive peak detection			UR	Upper right
* FB	Stop frequency	KSD	A + B → A			US	μsec
GZ	GHz	KSe	Negative peak detection	KZ	kHz	UV	μV
			Sample detection	LF	Preset 0 — 2.5 GHz	VB	Video BW
						0 to 9	0 to 9
						.	Decimal point or period

OUTPUT COMMANDS					
DR	Read display and increment address	MA	Marker amplitude output	* O3	Output format ASCII parameter or instrument units
EE	Enable number entry	MF	Marker frequency output	O4	Output format one 8 bit binary byte
KS< 91 >	Output amplitude error	OA	Output active function	TA	Output trace A
KS< 94 >	Output code for harmonic number in binary	OL	Output learn string	TB	Output trace B
KS< 123 >	Display read binary	OT	Output display text	UR	Upper right recorder output
KS< 126 >	Output every nth value of trace	O1	Output format ASCII display units		
LL	Lower left recorder output	O2	Output format two 8 bit binary bytes		

DISPLAY INPUT COMMANDS					
* DA	Display address	GR	Graph	* PD	Pen down
DD	Display write	IB	Input trace B, binary	PR	Plot relative
DW	Write into display and increment address	KS< 39 >	Fast Binary DA . . DW	PS	Skip to next display page
* D1	Display size normal	KS< 125 >	Display write binary	PU	Pen up
D2	Display size full CRT	KS< 127 >	Display write binary	SW	Skip to next control instruction
D3	Display size expand	LB	Label	TS	Take sweep
		PA	Plot absolute		

SERVICE REQUEST COMMANDS					
R1	Allow only SRQ 140	SRQ	Command	Bit	Definition
R2	Allow SRQ 140 and 104	102	R4	1	units key pressed
* R3	Allow SRQ 140 and 110	102	KS< 43 >	1	frequency limit exceeded
R4	Allow SRQ 140 and 102	104	R2	2	end of sweep
KS< 43 >	Allow SRQ 140 and 102	110	R3	3	hardware broken
		140	all	5	illegal command
		1xx	—	6	universal HP-IB service

\* selected with instrument preset

For more information, call your local HP Sales Office or nearest Regional Office: **Eastern** (201) 265-5000; **Midwestern** (312) 255-9800; **Southern** (404) 955-1500; **Western** (213) 970-7500; **Canadian** (416) 678-9430. Ask the operator for instrument sales. Or write Hewlett-Packard, 1501 Page Mill Road, Palo Alto, CA 94304. In **Europe**: Hewlett-Packard S.A., 7, rue du Bois-du-Lan, P.O. Box, CH 1217 Meyrin 2, Geneva, Switzerland. In **Japan**: Yokogawa-Hewlett-Packard Ltd., 29-21, Takaido-Higashi 3-chome, Sugiyama-ku, Tokyo 168.